



DEPARTMENT OF HEALTH & HUMAN SERVICES

Centers for Disease Control  
and Prevention (CDC)

NIOSH Reference: TN-24284  
Mfr. Reference: BYD0007

National Institute for Occupational  
Safety and Health (NIOSH)  
National Personal Protective  
Technology Laboratory (NPPTL)  
626 Cochran Mill Road  
Pittsburgh, PA 15236-0070  
Phone: 412-386-4000  
Fax: 412-386-4051  
October 30, 2020

Yubo Guan  
Product Manager  
BYD Precision Manufacture Co. Ltd.  
NO. 3001 Baohe Road, Baolong Industrial Area  
Longgang, Shenzhen, Guangdong Province  
CHINA

Dear Yubo Guan:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed the request referenced above and accepted October 1, 2020. This request was for extension of approval TC-84A-9221 for the model DE2322 N95 air-purifying filtering facepiece respirator. The user instructions are revised to change the expiration date from two to five years and the storage temperature as per recent validation reports. The complete respirator configuration is detailed on the assembly matrix, file name *0007AMa20201009.xlsx*, revision A/8, dated: September 20, 2020.

This request is granted. Approvals are granted only for documentation written in the English language. It is the manufacturer's responsibility to correctly translate materials desired in languages other than English.

Since this request does not affect the full approval labels, final labels are not enclosed.

The approved assembly consists of the parts as listed on the approval label and the assembly matrix. Parts are to be marked with the numbers indicated on the approval labels in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

Sincerely,

Jeffrey Peterson  
Chief, Conformity Verification and  
Standards Development Branch

Enclosure